



**HemaTechnologies**  
67 Beaver Ave  
Annandale, NJ 08801 (USA)

**Tel:** 908 823 9430  
**Fax:** 908 823 9428  
**E-mail:** sales@hematek.com

## ASSAY SHEET

### ESR CONTROL

Manufactured for HemaTechnologies by R&D Systems, Inc.

Lot # SR-0124

Expiration date: 05 July 2024

		LEVEL I Lot No. <b>SR 0124-1</b>		LEVEL II Lot No. <b>SR 0124-2</b>	
METHOD	UNITS	MEAN	RANGE	MEAN	RANGE
<b>ESR STAT™ PLUS</b>	mm/hr	9	4 - 14	68	38 - 98
<b>ESR STAT™ 6</b>	mm/hr	10	6 - 14	70	50 - 90

Note: 1. Follow the enclosed instructions for use.  
2. Assay values provided by HemaTechnologies

12/26/2023